

Spitalul Clinic „Prof.Dr. Th. Burghele”

Clinica Medicina Interna

SPECIFICAȚIE TEHNICĂ

Denumire echipament:SPIROMETRU..... -1 (UNA)..... buc.

Cerințe tehnice minimale	DA	NU	Conformitate
<p>Specificatii tehnice spirometru</p> <ul style="list-style-type: none">- Principiu de masurare: timp de tranzit prin ultrasunete- Parametric masurati: FVC, FVL, slow VC, MVV, Pre/Post, FVC, MVV6, FEV6, FEV1, FEV1/FVC, FEV1/VCmax, FEV1/FEV6, FEF25 (MEF75), FEF50 (MEF50), FEF75 (MEF25), FEF25-75% (MEF25-75%), PEF, FET, FIVC, PIF, ERV, IRV, Pre/Post % variation, lung age- Acuratetea masuratorilor: volume: $\pm 2\%$ or 0.050 l, flow $\pm 2\%$ or 0.020 l/s, PEF $\pm 5\%$ or 0.200 l/s, MVV $\pm 5\%$ or 5 l/min- Memorare minimum 100 teste- Certificare: CE Declaration of Conformity			
Instalare și punere în funcțiune – dacă este cazul			